

# Setting Sheet


## 5-3. Monoamines for Tissue Sample

### Column & Electrode Selection

Working Electrode	WE-3G (Graphite)	
Gasket	GS-25	
Separation Column	SC-50DS ( 3.0 x 150mm)	
PreColumn	PC-04AC *	position A
PreColumn	PC-04AC	position B

\* Can be left out.

### HTEC-510 Setting



### HPLC Line & Column Layout

